

N°2883983

**Calibration Register of Pulp Temperature Sensor**

<b>Container N°</b>	HLBU987662-1
---------------------	--------------

<b>Sensor N° 1</b>			
<b>Calibration</b>	<b>Reference Temperature</b>	<b>T° Sensor (°C)</b>	<b>Correction Factor Sensor N° 1</b>
1	0.0	0.1	-0.1
2	0.0	0.1	
3	0.0	0.1	
<b>Sensor N° 2</b>			
<b>Calibration</b>	<b>Reference Temperature</b>	<b>T° Sensor (°C)</b>	<b>Correction Factor Sensor N° 2</b>
1	0.0	0.0	0.0
2	0.0	0.0	
3	0.0	0.0	
<b>Sensor N° 3</b>			
<b>Calibration</b>	<b>Reference Temperature</b>	<b>T° Sensor (°C)</b>	<b>Correction Factor Sensor N° 3</b>
1	0.0	0.0	0.0
2	0.0	0.0	
3	0.0	0.0	

- The Correction Factor values were entered into the recorder and temperature Values in the treatment chart will be automatically adjusted.
- The Correction Factor values were NOT entered into the recorder and temperature Values in the treatment chart must be manually adjusted using the offset values.

*ABU*

ALBERTINA BENITEZ V.  
Ingeniera Agrónomo

Name, Signature and Stamp  
Authorized SAG Officer

