

N°2883982

Calibration Register of Pulp Temperature Sensor

Container N°	HLBU 629371-8
---------------------	---------------

Sensor N° 1			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 1
1	0.0	-0.2	0.2
2	0.0	-0.2	
3	0.0	-0.2	
Sensor N° 2			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 2
1	0.0	-0.2	0.2
2	0.0	-0.2	
3	0.0	-0.2	
Sensor N° 3			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 3
1	0.0	-0.2	0.2
2	0.0	-0.2	
3	0.0	-0.2	

- The Correction Factor values were entered into the recorder and temperature Values in the treatment chart will be automatically adjusted.
- The Correction Factor values were NOT entered into the recorder and temperature Values in the treatment chart must be manually adjusted using the offset values.

aby

ALBERTINA BENITEZ V.
Ingeniera Agrónomo

Name, Signature and Stamp
Authorized SAG Officer

