



Calibration Register of Pulp Temperature Sensor

Container N°	OTPU 666880-5
---------------------	----------------------

Sensor N° 1			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 1
1	0.0	-0.1	0.1
2	0.0	-0.1	
3	0.0	-0.1	
Sensor N° 2			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 2
1	0.0	-0.2	0.2
2	0.0	-0.2	
3	0.0	-0.2	
Sensor N° 3			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 3
1	0.0	-0.3	0.3
2	0.0	-0.3	
3	0.0	-0.3	

The Correction Factor values were entered into the recorder and temperature Values in the treatment chart will be automatically adjusted.

The Correction Factor values were NOT entered into the recorder and temperature Values in the treatment chart must be manually adjusted using the offset values.

SALIM HAMED GALDAMES

Name, Signature and Stamp
Authorized SAG Officer

