

N°2912343

Calibration Register of Pulp Temperature Sensor

Container N°	HLBU 939073-2
---------------------	----------------------

Sensor N° 1			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 1
1	0.0	0.1	-0.1
2	0.0	0.1	
3	0.0	0.1	
Sensor N° 2			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 2
1	0.0	0.1	-0.1
2	0.0	0.1	
3	0.0	0.1	
Sensor N° 3			
Calibration	Reference Temperature	T° Sensor (°C)	Correction Factor Sensor N° 3
1	0.0	0.1	-0.1
2	0.0	0.1	
3	0.0	0.1	

- The Correction Factor values were entered into the recorder and temperature Values in the treatment chart will be automatically adjusted.
- The Correction Factor values were NOT entered into the recorder and temperature Values in the treatment chart must be manually adjusted using the offset values.

aby

ALBERTINA BENITEZ V.
Ingeniera Agrónomo



Name, Signature and Stamp
Authorized SAG Officer